

DiaSys Diagnostic Systems GmbH
65558 Holzheim, GERMANY

Quality Control

Certificate of Analysis

QC release date: 2019-06-14

Product: TruLab CRP hs Level 1
Product Code: 59730

Cat. No.: 5 9730 99 10 046

Delivery of kits: No of kits: 5
Kit Lot/Los: 60137939
Kit Expiry: 2021-05

Results for:

Control Lot No.: 27210
Expiry: 2021-05

| Assay value | Range | Units |
|--------------------|--------------|--------------|
|--------------------|--------------|--------------|

| | | |
|-------|--------------|------|
| 0.953 | 0.696 - 1.21 | mg/L |
|-------|--------------|------|

Description: This Control is manufactured from a human serum base and all the donors are individually tested and shown using FDA-approved methods to be free from HBsAG, HCV and HIV antibodies.

Additional remarks: Mf-Date: 28.05.2019

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka